



In-house model verification for dsDNA

Sample Concentration (ng/ $\mu$ L)	Measurement Result (ng/ $\mu$ L)	Deviation(%)
5	5.4375	+8.8
50	52.395	+4.8
75	71.900	-4.1
300	307.83	+2.6
500	508.25	+1.7
900	883.10	-1.9
1500	1487.55	-0.8
2100	2123.48	+1.1